

Answer

Q1: **a** is correct. The image of the left side (**a**) is obtained from an ME collimator.

Background activity is lower due to lower septal penetration or scatter by the ME, compared with the LE collimator.